goodwork®
INTERNATIONAL

MASCARILLA KN95

MATERIAL SANITARIO

KN95

IMPERMEABLE, TRANSPIRABLE Y A PRUEBA DE FUEGO

Nos ayuda a protegernos contra agentes biológicos patógenos como virus o bactérias, así como aerosoles y líquidos no tóxicos como las neblinas de aceite.

goodwork®
INTERNATIONAL

MASCARILLA KN95
MATERIAL SANITARIO

FILTRO DE 5 CAPAS

Prevención eficaz frente al virus



ALIENTO	EFECTIVIDAD	FORMALDEHÍDO
<80 pa	>95 %	0 %

Características



Modelo: KN95

Tipo: No médica. NO estéril

Color: Blanco

Stándares de calidad: GB2626-2006 EN149:2001+A1:2009

Tamaño: 125 x 170 mm

Principales materiales: 5 capas. 29%-50 gr Tejido no-tejido. 14%-25 gr Tejido fundido. 14%-25 gr Tejido fundido. 29%-50 gr ES. 14%-25 gr Tejido no-tejido.

Propiedades de los productos: 95% de eficiencia de filtrado contra aerosoles sólidos y líquidos que no contienen aceites.

Periodo de almacenamiento: 2 años

goodwork®
INTERNATIONAL

MASCARILLA KN95

MATERIAL SANITARIO



Certificados

Las mascarillas autofiltrantes (recomendadas para personal sanitario con contacto menor a 2m) deben cumplir la norma UNE-EN 149:2001*A1:2009, según el Ministerio de Sanidad.

UNIVERSAL

UNIVERSAL
CERTIFICATION

NB 2163

EU TYPE EXAMINATION CERTIFICATE

Certificate No: 2163 - PPE - 667

Respiratory protective devices, filtering half masks to protect against particles manufactured for

DISTRIBUZIONE JUNIOR SRL
Via Pace, 25/26 - 80047 San Giuseppe Vesuviano (Na) ITALY
manufactured at
MEIZHUANGCHEN HEALTH TECHNOLOGY (SHENZHEN) CO., LTD
Meizhuangchen Health Technology (Shenzhen) Co. Ltd
No.12, Yuhe Road, Shiyan Town, Baoan District, Shenzhen, China.

are tested and evaluated according to

EN 149:2001 + A1:2009 Respiratory Protective Devices - Filtering Half Masks to Protect Against Particles - Requirements, Testing, Marking

Based on the type examination conducted with the evaluation of test reports, technical file according to Personal Protective Equipment Regulation (EU) 2016/425 Annex 5, it is approved that the product meets the requirements of the regulation.

Product Definition

Branda : ENHANCE **Model:** ENKN95-001
Filtering half mask
Total Inwards Leakage: Class – FFP2

Here by the manufacturer is allowed to use notified body number (2163) and can fix CE mark, as shown below, on the Category III product models given above, with;

- Issuing an appropriate EU Declaration of Conformity according to **Personal Protective Equipment Regulation (EU) 2016/425 Annex 9**.
- Ongoing successful performance in fulfilment of the requirements set out in Personal Protective **Equipment Regulation (EU) 2016/425** and harmonised standards, ensured by assessments based on **Annex 7 (Module C2) or Annex 8 (Module D)** of the regulation no later than 1 year from the beginning of serial production

This certificate is initially issued on **08 / 05 /2020** and will be valid for 5 years, if there is no change in the relevant harmonised standard affecting the essential health and safety requirements.

CE
2163

Suat KAÇMAZ
UNIVERSAL CERTIFICATION
Director

Verify the validity with the QR

Necip Fazıl Bulvarı Keyap Sitesi E2 Blok No:44/84 Yukarı Dudullu Ümraniye - İSTANBUL - TURKEY T:+90 216 455 80 80
UNIVERSALCERT.COM

goodwork®
INTERNATIONAL

MASCARILLA KN95
MATERIAL SANITARIO

UNIVERSAL

UNIVERSAL
CERTIFICATION

NB 2163

EU TYPE EXAMINATION CERTIFICATE

Certificate No: 2163 - PPE - 667

Respiratory protective devices, filtering half masks to protect against particles manufactured for

DISTRIBUZIONE JUNIOR SRL
Via Pace, 25/26 - 80047 San Giuseppe Vesuviano (Na) ITALY
manufactured at
MEIZHUANGCHEN HEALTH TECHNOLOGY (SHENZHEN) CO., LTD
Meizhuangchen Health Technology (Shenzhen) Co. Ltd
No.12, Yuhe Road, Shiyan Town, Baoan District, Shenzhen, China.

are tested and evaluated according to

EN 149:2001 + A1:2009 Respiratory Protective Devices - Filtering Half Masks to Protect Against Particles - Requirements, Testing, Marking

Based on the type examination conducted with the evaluation of test reports, technical file according to Personal Protective Equipment Regulation (EU) 2016/425 Annex 5, it is approved that the product meets the requirements of the regulation.

Product Definition

Branda : ENHANCE **Model:** ENKN95-001
Filtering half mask
Total Inwards Leakage: Class – FFP2

Here by the manufacturer is allowed to use notified body number (2163) and can fix CE mark, as shown below, on the Category III product models given above, with;

- Issuing an appropriate EU Declaration of Conformity according to **Personal Protective Equipment Regulation (EU) 2016/425 Annex 9**.
- Ongoing successful performance in fulfilment of the requirements set out in Personal Protective **Equipment Regulation (EU) 2016/425** and harmonised standards, ensured by assessments based on **Annex 7 (Module C2) or Annex 8 (Module D)** of the regulation no later than 1 year from the beginning of serial production

This certificate is initially issued on **08 / 05 /2020** and will be valid for 5 years, if there is no change in the relevant harmonised standard affecting the essential health and safety requirements.

CE
2163

Suat KAÇMAZ
UNIVERSAL CERTIFICATION
Director

Verify the validity with the QR

Necip Fazıl Bulvarı Keyap Sitesi E2 Blok No:44/84 Yukarı Dudullu Ümraniye - İSTANBUL - TURKEY T:+90 216 455 80 80
UNIVERSALCERT.COM

goodwork® INTERNATIONAL

MASCARILLA KN95
MATERIAL SANITARIO

Instrucciones de uso

1 - Saque el repirador del paquete y estírelo.

2 - Colocar la pinza nasal hacia arriba, el lado sin pinza hacia la cara, y estirar de las gomas para las orjas con ambas manos.

3 - Coloca la máscara contra la barbilla primero, y estira las gomas de las orejas y sujetalas hasta que te sientas cómodo.

4 - Ajusta el repirador y moldea la pinza nasal alrededor de la nariz para lograr un sellado seguro.

goodwork® INTERNATIONAL

Contacto

PARQUE EMPRESARIAL TÁCTICA
C/ Carboner 17 46980 Paterna, Valencia, España, Tel.(+34) 96 344 34 44

www.goodworkint.com